



Column: BTX5, 60m, 0.32 mm, 5µm, 120 kPa Helium
 Temp. program: 40°C / 10 min > 3°C/min > 295 °C / 40 min (135 min)
 Inj. Vol: 1µl; Split: 1:8
 EPED: 100 nm, 4 Hz, 50 ms, O₂: 20 kpa, H₂: 30 kpa
 Analog out: Cl*3, Multiplier 1

No.	Substance	(ng Cl)	No.	Substance	(ng Cl)
1	Methylenchloride	12,60	7	Chlorobenzene	3,96
2	Chloroforme	15,04	8	1,2,3-Trichloropropane	11,40
3	1,2-Dichloroethane	10,07	9	2-Chlorotoluene	3,45
4	1,1,1-Trichloroethane	12,14	10	4-Chlorotoluene	3,41
5	Carbontetrachloride	16,72	11	1,3-Dichlorobenzene	7,07
6	Tetrachloroethene	15,79			